

GenTrack Library Preparation Kit Certificate of Analysis

Product Information

Product Name: GenTrack Library Preparation Kit, 12 rxn

Product Catalog Number: NGS01201
Product Lot Number: 2019102901

Expiration Date: 01 Jul 2020

Storage: Store different reagents at -20°C or 4°C

Quality Control

Functional performance Tes	t*	Result
DNA library yield		Pass
Library size distribution		Pass
Mapping ratio		Pass
GC bias		Pass

^{*}The functional performance test is assessed in a library preparation experiment using NA12878 (*Coriell Institute*) genomic DNA and GA501 & GA701 from the GenTrack Adapter Set following the kit's minimum and maximum DNA input recommendations, respectively. Sequencing is performed on a NextSeq 550AR System (Illumina) to generate PE150 reads.

Caution

For Research use only. If you have any questions about the Certificate of Analysis, please contact our customer service representatives at 400-025-58897288-5812/5815, or Email: oligo@genscript.com for CN and oligo@genscript.com for International.

Version 2019-08-1

Approved by: Liuyufang

Quality release date: 06 Nov 2019