

Sample Name :Glucagon-Like Peptide (GLP) I(7-37)
 Cat No. :RP12738
 Lot No. :E7809921G
 Time Processed : 4:32:07 PM
 Month-Day-Year Processed : 08/07/2021

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow:1 ml/min
 Wavelength:220 nm

Time	Module	Command	Value
0.01	Pumps	Pump A B.Conc	5
25.00	Pumps	Pump A B.Conc	65
25.01	Pumps	Pump A B.Conc	95
31.00	Pumps	Pump A B.Conc	95
31.01	Pumps	Pump A B.Conc	5
40.00	Pumps	Pump A B.Conc	5
40.01	Controller	Stop	


<<Column Performance>>

<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm

Equipment: GK11010017

mV



1 Detector A Channel 1 / 220nm

Peak Table

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area %
1	13.939	137821	11843	0.842
2	18.263	309958	61688	1.895
3	18.447	15551623	1501705	95.061
4	19.233	83157	10620	0.508
5	20.881	259360	23375	1.585
6	23.528	17622	2358	0.108
Total		16359541	1611589	100.000